

LifePack 1000 AUTOMATED EXTERNAL DEFIBRILLATOR (3 LEAD) (CFD PARAMEDICS ONLY)

INDICATIONS

- AED 1000 is to be brought to patients' side for all EMS responses by CFD ALS ambulances.

CONTRAINDICATIONS

- None

EQUIPMENT

- LifePack AED 1000 with case
- 3 Lead cable
- 2 sets of defibrillation pads
- Electrodes (adult and pediatric)

PROCEDURE

1. Initiation of Patient Care policy B-2.
2. For known cardiac or respiratory calls, the manual monitor/defibrillator will be carried into patient's side.
3. The "3 Lead cable", with electrodes will be applied when indicated i.e. based on patient's history and/or chief complaint.
4. The defibrillation pads will be applied to the patient when indicated for defibrillation.
5. At anytime the LifePack 1000 (3-lead AED) is used, either for monitoring or for treatment, the patient will be immediately transferred to the cardiac monitor/defibrillator, once the patient has been transferred into the ambulance.

Copyright 2019 Chicago EMS Medical Directors Consortium

Written: 6/11

Reviewed: 6/11; 9/18

Revised: 9/18

MDC Approval: 6/16/11; 11/8/18

IDPH Approval: 9/29/11; 12/18/18

Implementation: 4/1/12; 1/18/19